


PLUTO Occipito-Cervical Fusion System

The PLUTO™ Occipito-Cervical Fusion System is a versatile system that is designed to be effective in treating the more challenging cases in the posterior occipito-cervical and upper thoracic spine.

The PLUTO™ System is comprised of cervical laminar hooks, thoracic poly-axial screws, and lateral rod connectors and midline occipital plates. The laminar hook is designed to provide excellent fit and fixation to the sub-axial cervical and upper thoracic spine. The poly-axial screw offers a degree of angulation and independent screw placement that reduces the need to contour the rod. A lateral rod connector provides a way to connect non-linear poly-axial screws to the rod. This enables us to best fit the anatomy of our patient. The occipital plate is intended to provide stabilization and promote fusion of the occipito-cervical junction.



Features & Benefits

- The implants provide the flexibility required to accommodate variations in patient anatomy.
- Variable angle self-tapping cortical and cancellous screws.
- System includes a complete set of implants and instruments designed to optimize fixation to the occiput.



Indications for Use

The PLUTO™ Spine System is designed for posterior stabilization of the occipito-cervical spine and upper thoracic spine. The implants provide the flexibility required to accommodate variations in patient anatomy.

- ◆ Instabilities in the occipito-cervical spine.
- ◆ Instabilities in the lower cervical and upper thoracic spine.
- ◆ Degenerative and painful post-traumatic conditions in the (insert space) lower cervical and upper thoracic spine.
- ◆ Anterior cervical fusions requiring additional posterior stabilization.

Implant Information [Material in Titanium Alloy (Ti-6Al-4V)]



Hook, 3.2mm

SIZE	REF
3.2 X 6.0 mm	MOI 37103060
3.2 X 4.5 mm	MOI 37103045



Crosslink Plate

SIZE	REF
25 - 30 mm	MOI 37104050
30 - 40 mm	MOI 37104060
40 - 60 mm	MOI 37104070



Double Rod Connector, 3.2mm

SIZE	REF
3.2 X 5.5 mm	MOI 37105055
3.2 X 6.0 mm	MOI 37105060



Lateral Rod Connector, 3.2mm

SIZE	REF
3.2 X 3.2 mm	MOI 37106060



Set Screw

SIZE	REF
3.2 mm	MOI 37107032



Occipital Screw

SIZE	REF
4.0 X 5.0 mm	MOI 37110050
4.0 X 6.0 mm	MOI 37110060
4.0 X 7.0 mm	MOI 37110070
4.0 X 8.0 mm	MOI 37110080



Midline Occipital Plate _ 3 Holes

SIZE	REF
3 Holes	MOI 37108003



Occipital Cervical Plate _ 4 Holes

SIZE	REF
4 Holes	MOI 37109004



Poly-Axial Screw (PAS), 3.5mm

SIZE	REF
3.5 X 12 mm	MOI 37101012
3.5 X 14 mm	MOI 37101014
3.5 X 16 mm	MOI 37101016
3.5 X 18 mm	MOI 37101018
3.5 X 20 mm	MOI 37101020
3.5 X 22 mm	MOI 37101022
3.5 X 24 mm	MOI 37101024
3.5 X 26 mm	MOI 37101026
3.5 X 28 mm	MOI 37101028
3.5 X 30 mm	MOI 37101030



Poly-Axial Screw (PAS), 4.0mm

SIZE	REF
4.0 X 12 mm	MOI 37102012
4.0 X 14 mm	MOI 37102014
4.0 X 16 mm	MOI 37102016
4.0 X 18 mm	MOI 37102018
4.0 X 20 mm	MOI 37102020
4.0 X 22 mm	MOI 37102022
4.0 X 24 mm	MOI 37102024
4.0 X 26 mm	MOI 37102026
4.0 X 28 mm	MOI 37102028
4.0 X 30 mm	MOI 37102030



Rod, 3.2mm

SIZE	REF
3.2 X 25 mm	MOI 37112025
3.2 X 30 mm	MOI 37112030
3.2 X 40 mm	MOI 37112040
3.2 X 50 mm	MOI 37112050
3.2 X 60 mm	MOI 37112060
3.2 X 70 mm	MOI 37112070
3.2 X 80 mm	MOI 37112080
3.2 X 90 mm	MOI 37112090
3.2 X 100 mm	MOI 37112100
3.2 X 200 mm	MOI 37112200

Occipital Plate / Rod (optional)

SIZE	REF
3.2 X 100 mm	MOI 37111100
3.2 X 200 mm	MOI 37111200





USA OFFICE

4000 Island Blvd, Suite 2704
Aventura, Florida 33160

PUERTO RICO OFFICE

1660 Santa Ana Street
Esq. Fidalgo Diaz,
San Juan, Puerto Rico 00909

E-mail:

info@madisonorthoinc.com

Website:

www.madisonorthoinc.com

